

Some Interpretations on FTIR Results for the Detection of *Ganoderma Boninense* in Oil Palm Tissue

¹Arnnyitte Alexander, ¹Jedol Dayou, ²Coswald Stephen Sipaut, ³ Chong Khim Phin and ⁴ Lee Ping Chin

¹Universiti Malaysia Sabah, School of Science and Technology, Energy, Vibration and Sound Research Group (e-VIBS), 88400 Kota Kinabalu, Sabah, Malaysia

²Universiti Malaysia Sabah, School of Engineering and Information Technology, Chemical Engineering Program, 88400 Kota Kinabalu, Sabah, Malaysia

³Universiti Malaysia Sabah, School of Science and Technology, Sustainable Palm Oil Research Unit (SPOR), 88400 Kota Kinabalu, Sabah, Malaysia

 4 Universiti Malaysia Sabah, School of Science and Technology, Biotechnology Programme, 88400 Kota Kinabalu, Sabah, Malaysia

ARTICLE INFO Article history: Received 25 June 2014 Received in revised form 8 July 2014 Accepted 14 September 2014 Available online 27 September 2014

Keywords:

Ganoderma boninense, oil palm, Fourier transform infrared spectroscopy, basal stem rot ABSTRACT Background: The use of Fourier Transform Infrared (FTIR) spectroscopy has been employed for the past three decades in the study of biological samples and characterization of biomolecules. **Objective:** In this paper, the author have conducted the feasibility study of detecting *Ganoderma boninense* in infected oil palm tree using Fourier transform infrared spectroscopy (FTIR). **Results:** It was found that result from FTIR plot is capable of showing the presence of *Ganoderma boninense* with good sensitivity. **Conclusion:** In this paper, the mechanisms behind this detection are discussed. This is due to the unique functional groups that exist in the *Ganoderma boninense* which cannot be found in healthy oil palm tissue.

© 2014 AENSI Publisher All rights reserved.

To Cite This Article: Arnnyitte Alexander, Jedol Dayou, Coswald Stephen Sipaut, Chong Khim Phin and Lee Ping Chin., Some Interpretations on FTIR Results for the Detection of Ganoderma Boninense in Oil Palm Tissue. Adv. Environ. Biol., 8(14), 30-32, 2014

INTRODUCTION

Ganoderma boninense is the causal pathogen of Basal Stem Rot (BSR) disease in oil palm. This disease is one of the most significant constraints to oil palm industry in the Southeast Asian countries. BSR was usually found on mature stands, however, diseased palms as young as 12-24 months after planting have been observed [12]. Infection of this disease can cause numerous yield losses and ultimately will result in the destruction of basal tissues hence death of diseases palms. Many efforts have been attempted to eliminate this disease, but none have given promising result [3]. In most cases, it failed due to late detection of the disease. This is because development BSR always symptomless and visible symptoms only appear at the very late stage of the infection when more than half of the root tissues have decayed, leaving no chance for the infected palms to be cured. Therefore, a viable early detection method for this disease has yet to be found. Currently, few methods have been reported to be useful to detect diseased palms even before the expression of symptoms. These include work on enzyme-linked immunosorbent assay (ELISA) [9] and Polymerase Chain Reaction (PCR) and sequence homology with specific primers for *Ganoderma boninense* [4], [8], [13]. However, the primers were reported to cross react with other saprophytic fungi and non *boninense* spp. Detection of *Ganoderma*-infected tissues using Ganoderma Selective Media (GSM) [2] was also reported, however, this method are time-consuming and not economic for large-scale application. In addition to these techniques, the development of PODITOOTM tomography to detect decayed tissues due to G. boninense infection based on sound lines also have been attempted [10]. Another work conducted, on utilising an electronic nose (E-nose) also had been reported in detecting BSR disease based on odour profiles [1]. This method provides fast result on field, however, it is not specifically designed to detect infection by G. boninense alone, as it also generates odour profiles when palms are infected by other pathogens. Therefore, a fast, reliable and accurate detection and quantification of G. boninense is still on demand. In the present study, Fourier-transform infrared (FTIR) spectroscopy was used for the identification and discrimination of functional group in G. boninense which could possibly used as an

Corresponding Author: Jedol Dayou, Universiti Malaysia Sabah, Energy, Vibration and Sound Research Group (e-VIBS), School of Science and Technology, 88400 Kota Kinabalu, Sabah, Malaysia. Tel: 60-88-320000; Fax: +60-88-435324; E-mail: jed@ums.edu.my

Advances in Environmental Biology, 8(14) Special 2014, Pages: 30-32

indicator in BSR early detection. The detection and identification of microorganisms using FTIR spectroscopy techniques is promising as valuable tool because of its sensitivity, rapidity, low cost and simplicity [11].

Methodology:

Pure culture of *Ganoderma boninense* was obtained from Genetic Laboratory of School of Science and Technology, Universiti Malaysia Sabah. The identity of *G. boninense* had been identified and confirmed using molecular technique [4]. The pure culture was then subcultured and maintained at 25°C on Potato Dextrose Agar (PDA). For comparison, healthy oil palm trunk tissues were taken from oil palm plantation in Sandakan, Sabah, Malaysia. Collection of trunk tissues was done according to the procedure described by [5]. The healthy tissues were confirmed free from *Ganoderma*-infection based on ergosterol analysis [6] and *Ganoderma* Selective Media (GSM) [2].

The FTIR spectrum was acquired using a Perkin Elmer 2000 Series Instrument. The spectrum resolution was set at 4 cm⁻¹ and the scanning range was selected from 650 to 4000 cm⁻¹. A small amount (~ 100 mg) of dried and powdered sample was placed onto the FTIR sample holder and spectra were collected. Three independent replicate samples of either *G. boninense* or healthy trunk tissues were measured.

RESULT AND DISCUSSION

In this study, the pure cultured of mycelium *G. boninense* was examined using FTIR to find spectroscopic biomarkers for rapid detection and identification of its colonization in oil palm. The result presented in Figure 1 shows there is a unique spectrum of *G. boninense* compared to healthy oil palm trunk tissues. Significant spectral differences between *G. boninense* and healthy trunk tissue can be seen in the range of $3500-3200 \text{ cm}^{-1}$ (Region A), $1650-1390 \text{ cm}^{-1}$ (Region B) and $1250-1000 \text{ cm}^{-1}$ (Region C) respectively. These clear differences provide a preliminary indication of possible spectral parameters for detection of *G. boninense*. The result was also supported by the fact that fungi pathogens display typical infrared spectra that differ from spectra of substrate material such as plant fibers, which make it is possible to detect and identify such pathogens directly from the infected tissue [7].



Fig. 1: FTIR spectra at the region 650-4000 cm⁻¹ of *Ganoderma boninense* and healthy oil palm trunk tissues.

Table 1 summaries the FTIR spectrum functional group analysis of *Ganoderma* and healthy tissues. The result shows the presence of several functional groups in *Ganoderma* which are not detected in healthy tissue sample. The identified functional group are N-H, C=N, C=H and C-O-C. This finding was in agreement with previous study conducted by [1] which found CH₃, CN and C-O-C in *Ganoderma*'s fruiting bodies. Most of the functional groups in *Ganoderma* are detected in the area ranged between 1000 to 1800 cm⁻¹, which can be considered as an important area for easy and reliable discrimination with the healthy trunk tissue.

Wave-length (cm ⁻¹)	Region	Functional group		Possible Functional group for indicator of <i>Ganoderma</i> infection
		Ganoderma boninense	Healthy trunk tissue	
3500-3200	А		O-H (phenol)	
3400-3200	А	N-H (amine)		N-H (amine)
1650-1600	В	C=O (amide)	C=O (amide)	
1580-1500	В	C=N (imine)		C=N (imine)
1470-1450	В	C-H (alkane)		C-H (alkane)
1400-1390	В		C-O(carboxylic acid)	
1250-1000	С	C-O-C (ether)		C-O-C (ether)
1100-1000	С		Si-O (silicone)	

Table 1: Functional Group of G. boninense mycelia and healthy oil palm trunk tissues.

Empty box indicate absence of the respective compound in the sample(s).

Advances in Environmental Biology, 8(14) Special 2014, Pages: 30-32

Conclusion:

In conclusion, the present study shows that there is a great potential of FTIR spectroscopy for an easy and rapid detection of *G. boninense* infection in oil palm due to the specific functional groups that exist in the *Ganoderma boninense* which cannot be found in healthy oil palm tissue. Detection of BSR infection in oil palm at early stage is crucial for management of this disease.

ACKNOWLEDGEMENT

The work presented in this paper is supported the Malaysian Ministry of Education under Fundamental Research Grant Scheme no FRG0342-ST-2/2013, and is greatly acknowledged.

REFERENCES

- Abdullah, A.H., A.Y. Md. Shakaff, A.H. Adom, M.N. Ahmad, A. Zakaria, S.A. Ghani, N.M. Samsudin, F. S.A. Saad, L.M. Kamaruddin, N.H. Hamid and I.A. Seman. 2012. Exploring MIP Sensor of Basal Stem Rot (BSR) Disease in Palm Oil Plantation. In *The 14th International Meeting on Chemical Sensors*, Nurembery, Germany.
- [2] Ariffin, D. and A.S. Idris, 1992. The *Ganoderma* Selective Medium (GSM). *PORIM Information Series no.8.* Persiaran Institusi, Kajang, Malaysia: Palm Oil Research Institute of Malaysia.
- [3] Bivi, M.R., M.S.N. Farhana, A. Khairulmazmi and A. Idris, 2010. Control of *Ganoderma boninense*: A causal agent of basal stem rot disease in oil palm with endophyte bacteria *in vitro*. International Journal of Agriculture Biology, 12: 833-839.
- [4] Chong, K.P., C.P. Foong, C.M.V.L. Wong, S. Rossall and M. Atong. 2011. First identification of *Ganoderma boninense* isolated from Sabah based on PCR and sequence homology. African Journal of Biotechnology, 10(66): 14718-14723.
- [5] Chong, K.P. 2012. An evaluation of The *Ganoderma* Fungal Colonisation Using Ergosterol Analysis and Quantification. The Planter, Kuala Lumpur, 88(1034): 311-319.
- [6] Chong, K.P., A. Markus and S. Rossall. 2012. The susceptibility of different varieties of oil palm seedlings to *Ganoderma boninense* infection. Pakistan Journal of Botany, 44(6): 2001-2004.
- [7] Erukhimovitch, V., L. Tsror, Hazanovsky, M. Talyshinsky, I. Mukmanov, Y. Souprun and M. Huleihel. 2005. Identification of fungal phyto-pathogens by Fourier-transfrom infrared (FTIR) microscopy. Journal of Agricultural Technology, 1(1): 145-152. 2005.
- [8] Idris, A.S., M. Yamaoka, S. Hayakawa, M.W. Basri, I. Noorhasimah and D. Ariffin. 2003. PCR technique for detection of *Ganoderma*. MPOB Information. Series MPOB TT No. 188.
- [9] Idris, A.S. and A.R. Rafidah, 2008. Polyclonal antibody for detection of *Ganoderma*. MPOB Information. Series MPOB TT, 430.
- [10] Idris, A.S., M.S. Mazliham and A.Z. Madihah, 2009. Current technologies for detection of *Ganoderma* in oil palm. In (Unedited) *Proceedings of Agriculture, Biotechnology and Sustainability Conference*, PIPOC, Kuala Lumpur, 81-99.
- [11] Salman, A., L. Tsror, A. Pomerantz, R. Moreh, S. Mordechai and M. Huleihel, 2010. FTIR spectroscopy for detection and identification of fungal phytopathogens. Spectroscopy, 24: 261-267.
- [12] Singh, G., 1991. Ganoderma the scourge of oil palm in the coastal areas. The Planter, (67): 421-44
- [13] Yamaoka, M., S. Hayakawa, M. Tsukamoto, R. Kurane, A.S. Idris, H. Mohd Haniff and D. Ariffin, 2000. Diagnosis of basal stem rot of oil palm by foliar analysis and PCR-based detection of *Ganoderma* from tissues of the diseased oil palm. *In: 23rd Malaysian Society of Microbiology Symposium*, Langkawi, Kedah.